



Designation : Rk measurement
Testorder : RedFac
Test Date/Time : 6/12/2024 10:15:09 AM Sample-ID-No. : 000008X1YR

Start :

Test Result: PASS

Frequency: 50.0 [Hz]

Em [mV]	Ea [mV]	rk	rk Limit
2	0.2662	0.133	0.150
5	0.5781	0.116	0.150
10	0.9609	0.096	0.150
20	1.552	0.078	0.150
50	2.598	0.052	0.150
100	3.466	0.035	0.150
200	4.781	0.024	0.150
300	6.553	0.022	0.150
400	9.793	0.024	0.150
500	16.64	0.033	0.150
600	29.6	0.049	0.150
700	49.26	0.070	0.150
800	72.85	0.091	0.150
900	99.64	0.111	0.150
1000	128.4	0.129	0.150

